



AKKREDITOINTITODISTUS

ACCREDITATION CERTIFICATE

**VAISALA OYJ
MITTANORMAALILABORATORIO
VANTAA
K008**

on FINAS-akkreditointipalvelun akkreditoima kalibrointilaboratorio K008.
Akkreditointipäätös on voimassa 8.5.2013 asti.
Tiedot akkreditoinnin pätevyysalueesta ja voimassaolosta on esitetty
verkkosivulla www.finas.fi.

is calibration laboratory No. K008 accredited by FINAS Finnish Accreditation Service.
The accreditation decision is valid until 8.5.2013.
Information about the scope and the current status of the accreditation is available
on the web page www.finas.fi.

Toimielin täyttää seuraavan standardin vaatimukset:
The above body conforms to the requirements of the following standard:

SFS-EN ISO/IEC 17025:2005

Espoo 8.5.2009

A handwritten signature in black ink, appearing to read "Leena Tikkainen".
Leena Tikkainen

A handwritten signature in black ink, appearing to read "Seppo Lehto".
Seppo Lehto

CALIBRATION CERTIFICATE

Instrument HM34C Humidity and temperature meter
Serial number G0740002
Manufacturer Vaisala Oyj, Finland
Calibration date 17th February 2011
Test procedure

The above instrument was calibrated by comparing the relative humidity and temperature readings to two HMP233 factory working standards. At the time of shipment, the instrument described above met its operating specifications.

The relative humidity readings of the two HMP233 factory working standards have been calibrated at the Vaisala factory by using Hygro M-3 dewpoint meter. Hygro M-3 dewpoint meter has been calibrated at the NIST laboratories by using the NIST primary standard. The temperature readings of the two HMP233 factory working standards have been calibrated at Vaisala Measurement Standards Laboratory (MSL) by using the MSL working standard traceable to the NIST. MSL has been accredited by the FINAS according to the ISO/IEC 17025 (K008).

Calibration results

Reference humidity* %RH	Observed humidity %RH	Difference %RH	Permissible difference %RH
0.5	0.5	0.0	± 2.0
45.6	45.1	- 0.5	± 2.0
71.4	71.3	- 0.1	± 2.0
Reference temperature* °C	Observed temperature °C	Difference °C	Permissible difference °C
23.61	23.70	+ 0.09	0.2

*Average of two references

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
Vaisala HMT337	D2350020	2010-12-21	H06-10510070
Vaisala HMT337	D2350026	2010-12-21	H06-10510071
Vaisala HMT337	D2350020	2010-12-08	K008-T02443
Vaisala HMT337	D2350026	2010-12-08	K008-T02441
HYGRO M-3	361095	2010-07-07	M-10H039

Uncertainties (95 % confidence level, k=2)

Temperature ± 0.13 °C.
Ambient conditions / Humidity $36 \pm 5\%$ RH, Temperature 23 ± 1 °C, Pressure 1030 ± 1 hPa.

Technician

This report shall not be reproduced except in full, without the written approval of Vaisala.

Doc210425v97